

SETARI PENTRU APARATUL ARIAC 98

Apparatus Information

Manufacturer: Model:

Apparatus Type

- Compressed Air Breathing Apparatus
- Airline Apparatus
- Combination CABA/Airline
- Cylinder
- Hood
- Closed Circuit Breathing Apparatus
- Chemical Protection Suit

Supply Pressure

- 200 bar
- 300 bar

Apparatus Type | Apparatus Information | Leak, Alarm Test Parameters

Activation/Static Test | Breathing Test Parameters | Pressure Gauge Test

Apparatus Information

Manufacturer: Model:

First Breath Activation Pressure

- Yes
- No

Allow 40 x 2.5 L/min Work in Breathing Tests

- Yes
- No

Perform Medium Pressure Tests

- Yes
- No

Gauge Type

- Fraction/Percent
- Digital
- Numbers
- HUD with no other gauge

Alarm Type

- Bell
- Whistle
- LED

Redundant Low Pressure Alarm

- Yes
- No

High Pressure Hose Volume

Liters

Apparatus Type | **Apparatus Information** | Leak, Alarm Test Parameters

Activation/Static Test | Breathing Test Parameters | Pressure Gauge Test

SETARI PENTRU APARATUL ARIAC 98

Apparatus Information
✕

Manufacturer Model

High Pressure Leak Test

Test Duration Sec.

Pressure Change Allowed bar

Alarm Activation Pressure

Minimum bar

Maximum bar

Facepiece Leak Test

Starting Pressure mbar

Allowable Leakage mbar

Test Duration Sec.

Exhalation Valve Opening Pressure

Minimum mbar

Maximum mbar

Apparatus Type

Apparatus Information

Leak, Alarm Test Parameters

Activation/Static Test

Breathing Test Parameters

Pressure Gauge Test

Apparatus Information
✕

Manufacturer Model

Static Facepiece Pressure

Delay Minimum mbar

Maximum mbar

Medium Pressure Change

Maximum bar

Static Medium Pressure

Minimum bar

Maximum bar

Safety Valve Opening Pressure

Minimum bar

Maximum bar

Bypass Flow

Minimum L/min

Maximum L/min

Activation/Static Test

Breathing Test Parameters

Pressure Gauge Test

Apparatus Type

Apparatus Information

Leak, Alarm Test Parameters

SETARI PENTRU APARATUL ARIAC 98

Apparatus Information

Manufacturer: Model:

Breathing Resistance Facepiece Pressure

Breathing Rate	Minimum	Maximum
25 x 2 L/min Breathing Rate	-7 mbar	7 mbar
40 x 2.5 L/min Breathing Rate	-10 mbar	10 mbar

Dynamic Medium Pressure

Breathing Rate	Minimum	Maximum
25 x 2 L/min Breathing Rate	4 bar	9 bar
40 x 2.5 L/min Breathing Rate	3 bar	9.5 bar

Activation/Static Test **Breathing Test Parameters** Pressure Gauge Test

Apparatus Type Apparatus Information Leak, Alarm Test Parameters

Apparatus Information

Manufacturer: Model:

Pressure Gauge Test

Maximum Error	10 bar
Maximum Error at Test Point 3	5 bar

Pressure Gauge Test Points

Test Point 1 (Highest)	150 bar
Test Point 2	100 bar
Test Point 3 (Lowest)	70 bar

Activation/Static Test Breathing Test Parameters **Pressure Gauge Test**

Apparatus Type Apparatus Information Leak, Alarm Test Parameters

TASTA "OK" SE APASA DOAR DUPA CE S-AU INTRODUS SETARILE IN TOATE CELE 6 FERESTRE - PENTRU CONFIRMARE.